



# National Institute of Standards & Technology

## Certificate of Analysis

### Standard Reference Material<sup>®</sup> 2735

#### Nitric Oxide in Nitrogen

(Nominal Amount-of-Substance Fraction – 800  $\mu\text{mol/mol}$ )

*This certificate reports the certified values for Lot 141-C-XX.*

This Standard Reference Material (SRM) is a primary gas mixture that, the amount-of-substance fraction expressed as concentration [1], may be related to secondary working standards. The SRM is intended for the calibration of instruments used for nitric oxide determinations and for other uses.

This SRM mixture is supplied in a DOT 3AL specification aluminum (6061 alloy) cylinder with a water volume of 6 L. Mixtures are shipped with a nominal pressure exceeding 12.4 MPa (1800 psig), which provides the user with 0.73 m<sup>3</sup> (25.8 ft<sup>3</sup>) of useable mixture. The cylinder is the property of the purchaser and is equipped with a CGA-660 stainless steel valve, which is the recommended outlet for this nitric oxide mixture. NIST recommends that this cylinder **NOT** be used below 0.7 MPa (100 psig).

**Certified Value:** This SRM mixture has been certified for nitric oxide (NO) and total oxides of nitrogen (NO<sub>x</sub>) concentrations. The certified value, given below, applies to the identified cylinder and NIST sample number.

Nitric Oxide (NO) Concentration: 786.9  $\mu\text{mol/mol}$   $\pm$  1.7  $\mu\text{mol/mol}$

Total Oxides of Nitrogen (NO<sub>x</sub>) Concentration: 788.2  $\mu\text{mol/mol}$   $\pm$  1.9  $\mu\text{mol/mol}$

Cylinder Number:

NIST Sample Number:

The uncertainty of the certified value includes the estimated uncertainty in the NIST standards, the analytical comparisons to the lot standard (LS), and the uncertainty of comparing the LS with each of the mixtures comprising this lot. This uncertainty is expressed as an expanded uncertainty,  $U = ku_c$ , with  $u_c$  determined by experiment and a coverage factor  $k = 2$ . The true value for the nitric oxide amount-of-substance fraction is asserted to lie in the interval defined by the certified value  $\pm U$  with a level of confidence of approximately 95 % [2].

**Expiration of Certification:** This certification is valid until **01 February 2012**, within the measurement uncertainties specified, provided the SRM is handled and stored in accordance with the instructions given in this certificate. However, the certification will be nullified if the SRM is contaminated or modified. NIST will monitor this SRM over the period of its certification. If changes occur that affect the certification before the expiration of this certificate, NIST will notify the purchaser. Registration (see attached sheet) will facilitate notification.

**Hydrotest Date:** 08/04

**Blend Date:** 09/04

**Cylinder and Gas Handling Information:** NIST recommends the use of a high purity, stainless steel, two-stage pressure regulator with a stainless steel diaphragm and CGA-660 outlet to safely reduce the pressure and to deliver this SRM mixture to the instrument. The regulator should be purged and evacuated several times to prevent accidental contamination of the sample.

The analytical measurements leading to the certification of this current SRM lot were performed by M.E. Kelley of the NIST Analytical Chemistry Division.

Stephen A. Wise, Chief  
Analytical Chemistry Division

Gaithersburg, MD 20899  
Certificate Issue Date: 04 May 2006

Robert L. Watters, Jr., Chief  
Measurement Services Division

The overall direction and coordination of the technical work required for certification of this SRM lot were performed by F.R. Guenther of the NIST Analytical Chemistry Division.

The support aspects involved in the issuance of this SRM were coordinated through the NIST Measurement Services Division.

**Mixture Preparation:** The gas mixtures comprising this SRM lot were prepared in accordance with NIST technical specifications by a commercial specialty gas vendor under contract to NIST. The specifications stipulate that each SRM mixture be identical in nitric oxide concentration and stable with time.

**Analytical Methods:** Analyses of the nitric oxide concentration for this lot of cylinders were conducted by comparing each cylinder mixture to a representative cylinder chosen from the lot, the lot standard (LS), using chemiluminescence. Assignment of the nitric oxide concentration to the LS was accomplished by comparison to primary gravimetric standards using chemiluminescence.

**Homogeneity Analysis:** Each of the nitric oxide mixtures that comprise this SRM lot was compared to the LS using chemiluminescence. An analysis of variance indicated that sample-to-sample nitric oxide concentration differences were statistically insignificant within the lot. Therefore, a single concentration has been assigned to the entire SRM lot.

**Nitric Oxide Concentration Value Assignment:** The certified nitric oxide concentration for this SRM cylinder was computed from the assigned concentration for the lot standard and the homogeneity analysis.

**Total Oxides of Nitrogen Concentration Value Assignment:** The certified total oxides of nitrogen concentration for each SRM cylinder was computed from the assigned concentration for the lot standard and the homogeneity analysis.

**Stability:** Periodic analyses of SRM units from this lot are performed at NIST to monitor stability. If significant changes in the nitric oxide concentration are observed, the purchaser will be notified. Refer to the "Cylinder and Gas Handling Information" section for proper handling of this SRM.

## REFERENCES

- [1] Taylor, B.N.; *Guide for the Use of the International System of Units (SI)*; NIST Special Publication 811; U.S. Government Printing Office: Washington, DC (1995).
- [2] ISO; *Guide to the Expression of Uncertainty in Measurement*; ISBN 92-67-10188-9, 1st ed.; International Organization for Standardization: Geneva, Switzerland (1993); see also Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297; U.S. Government Printing Office: Washington, DC (1994); available at <http://physics.nist.gov/Pubs/>.

*Users of this SRM should ensure that the certificate in their possession is current. This can be accomplished by contacting the SRM Program at: telephone (301) 975-6776; fax (301) 926-4751; e-mail [srminfo@nist.gov](mailto:srminfo@nist.gov); or via the Internet at <http://www.nist.gov/srm>*